

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022322\
 Data File : VN071045.D
 Acq On : 23 Feb 2022 20:09
 Operator : JC\MD
 Sample : N1638-01 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30960

Quant Time: Feb 24 00:42:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	757962	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1213130	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1078274	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	360806	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	460291	53.162	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.320%
35) Dibromofluoromethane	8.022	113	354438	49.120	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.240%
50) Toluene-d8	10.439	98	1471589	50.784	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.560%
62) 4-Bromofluorobenzene	12.727	95	485835	46.905	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.800%
Target Compounds						
						Qvalue
16) Acetone	4.298	43	33687	7.065	ug/l #	59
25) 2-Butanone	7.345	43	8773	1.383	ug/l	91
40) Benzene	8.463	78	12039	0.344	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022322\
Data File : VN071045.D
Acq On : 23 Feb 2022 20:09
Operator : JC\MD
Sample : N1638-01 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30960

Quant Time: Feb 24 00:42:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

