

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022322\
 Data File : VN071043.D
 Acq On : 23 Feb 2022 19:21
 Operator : JC\MD
 Sample : N1642-06 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31458

Quant Time: Feb 24 00:41:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	876010	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1373782	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1230546	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	418554	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	523434	52.308	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.620%
35) Dibromofluoromethane	8.016	113	402636	49.275	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.540%
50) Toluene-d8	10.439	98	1676041	51.075	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.160%
62) 4-Bromofluorobenzene	12.727	95	558262	47.595	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.180%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022322\
Data File : VN071043.D
Acq On : 23 Feb 2022 19:21
Operator : JC\MD
Sample : N1642-06 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31458

Quant Time: Feb 24 00:41:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.086 min Scan# 1036
Delta R.T. 0.005 min
Lab File: VN071043.D
Acq: 23 Feb 2022 19:21




Instrument :
MSVOA_N
ClientSampleId :
31458


Tgt Ion:168 Resp: 876010
Ion Ratio Lower Upper
168 100
99 54.8 41.3 61.9



#33
1,2-Dichloroethane-d4
Concen: 52.308 ug/l
RT: 8.433 min Scan# 1096
Delta R.T. 0.006 min
Lab File: VN071043.D
Acq: 23 Feb 2022 19:21

Tgt Ion: 65 Resp: 523434
Ion Ratio Lower Upper
65 100
67 52.8 0.0 103.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.969 min Scan# 1186

Delta R.T. 0.006 min

Lab File: VN071043.D

Acq: 23 Feb 2022 19:21

Instrument :

MSVOA_N

ClientSampleId :

31458





Tgt Ion:114 Resp: 1373782

Ion Ratio Lower Upper

114 100

63 21.3 0.0 37.8

88 15.5 0.0 27.4



#35

Dibromofluoromethane

Concen: 49.275 ug/l

RT: 8.016 min Scan# 1025

Delta R.T. -0.000 min

Lab File: VN071043.D

Acq: 23 Feb 2022 19:21




Tgt Ion:113 Resp: 402636


Ion Ratio Lower Upper

113 100

111 103.6 82.2 123.2

192 20.6 16.9 25.3









#50
Toluene-d8
Concen: 51.075 ug/l
RT: 10.439 min Scan# 1436
Delta R.T. 0.006 min
Lab File: VN071043.D
Acq: 23 Feb 2022 19:21




Instrument :
MSVOA_N
ClientSampleId :
31458


Tgt Ion: 98 Resp: 1676041
Ion Ratio Lower Upper
98 100
100 66.1 50.4 75.6



#62
4-Bromofluorobenzene
Concen: 47.595 ug/l
RT: 12.727 min Scan# 1826
Delta R.T. 0.006 min
Lab File: VN071043.D
Acq: 23 Feb 2022 19:21

Tgt Ion: 95 Resp: 558262
Ion Ratio Lower Upper
95 100
174 80.7 0.0 187.2
176 77.1 0.0 181.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.739 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN071043.D
Acq: 23 Feb 2022 19:21

Instrument :
MSVOA_N
ClientSampleId :
31458

Tgt Ion:117 Resp: 1230546
Ion Ratio Lower Upper
117 100
82 58.1 42.4 63.6
119 32.1 25.4 38.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.674 min Scan# 1987
Delta R.T. 0.006 min
Lab File: VN071043.D
Acq: 23 Feb 2022 19:21

Tgt Ion:152 Resp: 418554
Ion Ratio Lower Upper
152 100
115 57.8 27.5 82.5
150 157.9 0.0 344.4

