

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022323\
 Data File : VN076789.D
 Acq On : 23 Feb 2023 13:22
 Operator : JC\MD
 Sample : 01664-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Feb 24 01:34:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	274063	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	489998	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	436149	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	163394	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	196358	51.886	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.780%	
35) Dibromofluoromethane	8.172	113	152938	47.660	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.320%	
50) Toluene-d8	10.571	98	591699	49.375	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.740%	
62) 4-Bromofluorobenzene	12.854	95	173449	43.700	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 87.400%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022323\
Data File : VN076789.D
Acq On : 23 Feb 2023 13:22
Operator : JC\MD
Sample : 01664-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Feb 24 01:34:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

