

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022323\
 Data File : VN076792.D
 Acq On : 23 Feb 2023 14:49
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Feb 24 01:35:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	275540	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	512701	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	455243	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	184680	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	199198	52.354	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.700%
35) Dibromofluoromethane	8.178	113	153193	45.626	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.260%
50) Toluene-d8	10.572	98	622536	49.648	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.300%
62) 4-Bromofluorobenzene	12.854	95	188848	45.473	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022323\
Data File : VN076792.D
Acq On : 23 Feb 2023 14:49
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Feb 24 01:35:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

