

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022324\
 Data File : VN081163.D
 Acq On : 23 Feb 2024 16:56
 Operator : JC\MD
 Sample : VN0223WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0223WBL02

Quant Time: Feb 23 22:00:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	286867	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	538714	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	476375	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	194073	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	179445	51.034	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.060%
35) Dibromofluoromethane	8.171	113	149773	49.166	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.340%
50) Toluene-d8	10.571	98	474872	43.890	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	87.780%
62) 4-Bromofluorobenzene	12.853	95	180776	44.466	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022324\
Data File : VN081163.D
Acq On : 23 Feb 2024 16:56
Operator : JC\MD
Sample : VN0223WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0223WBL02

Quant Time: Feb 23 22:00:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

