

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022324\
 Data File : VN081170.D
 Acq On : 23 Feb 2024 19:45
 Operator : JC\MD
 Sample : PB159217TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB159217TB

Quant Time: Feb 23 22:04:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	266381	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	513995	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	470160	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	183584	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	186864	57.231	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.460%
35) Dibromofluoromethane	8.171	113	156731	53.925	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.840%
50) Toluene-d8	10.571	98	516978	50.080	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.160%
62) 4-Bromofluorobenzene	12.853	95	186520	48.086	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.180%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	62214	48.425	ug/l	90
20) Methylene Chloride	5.283	84	26576	7.200	ug/l	90
37) Ethyl Acetate	7.571	43	7573	1.685	ug/l #	95
43) Isopropyl Acetate	8.694	43	464453	59.093	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022324\
Data File : VN081170.D
Acq On : 23 Feb 2024 19:45
Operator : JC\MD
Sample : PB159217TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB159217TB

Quant Time: Feb 23 22:04:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

