

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022324\
 Data File : VN081173.D
 Acq On : 23 Feb 2024 20:57
 Operator : JC\MD
 Sample : P1574-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-3

Quant Time: Feb 23 22:05:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	274872	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	547499	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	500764	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	194185	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	178090	52.859	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.720%
35) Dibromofluoromethane	8.171	113	148114	47.841	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.680%
50) Toluene-d8	10.571	98	487938	44.374	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.740%
62) 4-Bromofluorobenzene	12.853	95	181428	43.911	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.820%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	68538	51.700	ug/l	94
20) Methylene Chloride	5.289	84	18672	4.600	ug/l	88
37) Ethyl Acetate	7.565	43	7212	1.507	ug/l #	89
43) Isopropyl Acetate	8.694	43	469416	56.070	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022324\
Data File : VN081173.D
Acq On : 23 Feb 2024 20:57
Operator : JC\MD
Sample : P1574-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-3

Quant Time: Feb 23 22:05:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

