

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022324\
 Data File : VN081175.D
 Acq On : 23 Feb 2024 21:45
 Operator : JC\MD
 Sample : P1574-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-5

Quant Time: Feb 24 01:11:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	265146	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	532812	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	479388	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	184679	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	182570	56.176	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.360%
35) Dibromofluoromethane	8.171	113	152571	50.639	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.280%
50) Toluene-d8	10.571	98	486089	45.425	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.840%
62) 4-Bromofluorobenzene	12.853	95	175088	43.544	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.080%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	69834	54.610	ug/l	95
18) Methyl Acetate	5.030	43	7463	2.392	ug/l	98
20) Methylene Chloride	5.277	84	19889	5.179	ug/l	90
43) Isopropyl Acetate	8.688	43	474512	58.241	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022324\
Data File : VN081175.D
Acq On : 23 Feb 2024 21:45
Operator : JC\MD
Sample : P1574-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-5

Quant Time: Feb 24 01:11:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

