

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022324\
 Data File : VN081176.D
 Acq On : 23 Feb 2024 22:09
 Operator : JC\MD
 Sample : P1574-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-6

Quant Time: Feb 24 01:11:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	279141	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	546519	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	496145	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	195836	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	175140	51.188	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.380%
35) Dibromofluoromethane	8.171	113	149351	48.327	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.660%
50) Toluene-d8	10.571	98	486881	44.358	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.720%
62) 4-Bromofluorobenzene	12.853	95	177564	43.052	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.100%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	69293	51.470	ug/l	95
18) Methyl Acetate	5.030	43	7099	2.162	ug/l #	82
20) Methylene Chloride	5.283	84	19785	4.841	ug/l #	82
37) Ethyl Acetate	7.559	43	8803	1.842	ug/l	96
43) Isopropyl Acetate	8.688	43	572637	68.522	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022324\
Data File : VN081176.D
Acq On : 23 Feb 2024 22:09
Operator : JC\MD
Sample : P1574-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-6

Quant Time: Feb 24 01:11:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

