

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022420\
 Data File : VN060202.D
 Acq On : 24 Feb 2020 12:15
 Operator : JC/MD
 Sample : PB127002TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127002TB

Quant Time: Feb 24 16:22:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	189832	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	300007	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	285946	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	127706	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	111548	48.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.58%	
35) Dibromofluoromethane	7.57	113	91106	49.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.80%	
50) Toluene-d8	10.08	98	364016	50.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.26%	
62) 4-Bromofluorobenzene	12.40	95	132990	49.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.68%	
Target Compounds						
16) Acetone	3.79	43	6931	6.879	ug/l	97
20) Methylene Chloride	4.53	84	9681	4.816	ug/l	95
43) Isopropyl Acetate	8.14	43	42225	9.836	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022420\
Data File : VN060202.D
Acq On : 24 Feb 2020 12:15
Operator : JC/MD
Sample : PB127002TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127002TB

Quant Time: Feb 24 16:22:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

