

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022420\
 Data File : VN060207.D
 Acq On : 24 Feb 2020 14:09
 Operator : JC/MD
 Sample : L1659-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WH-2-COMP-2

Quant Time: Feb 24 16:51:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	181340	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	280809	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	267212	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	119938	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	104538	47.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.76%	
35) Dibromofluoromethane	7.57	113	84965	49.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.44%	
50) Toluene-d8	10.08	98	341933	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
62) 4-Bromofluorobenzene	12.40	95	124268	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
Target Compounds						
16) Acetone	3.80	43	3311	3.440	ug/l	93
20) Methylene Chloride	4.53	84	7049	3.671	ug/l	92
43) Isopropyl Acetate	8.14	43	42921	10.681	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022420\
Data File : VN060207.D
Acq On : 24 Feb 2020 14:09
Operator : JC/MD
Sample : L1659-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WH-2-COMP-2

Quant Time: Feb 24 16:51:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

