

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022420\
 Data File : VN060210.D
 Acq On : 24 Feb 2020 15:17
 Operator : JC/MD
 Sample : PB127001ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127001ZHE#03

Quant Time: Feb 24 17:00:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	181927	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	285907	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	272396	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	123589	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	105844	47.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.62%	
35) Dibromofluoromethane	7.57	113	86305	49.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.22%	
50) Toluene-d8	10.08	98	351367	50.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.54%	
62) 4-Bromofluorobenzene	12.40	95	127513	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
Target Compounds						
16) Acetone	3.79	43	4333	4.487	ug/l	99
20) Methylene Chloride	4.53	84	8677	4.504	ug/l	94
43) Isopropyl Acetate	8.14	43	42580	10.408	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022420\
Data File : VN060210.D
Acq On : 24 Feb 2020 15:17
Operator : JC/MD
Sample : PB127001ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127001ZHE#03

Quant Time: Feb 24 17:00:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

