

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022420\
 Data File : VN060211.D
 Acq On : 24 Feb 2020 15:40
 Operator : JC/MD
 Sample : PB127001ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127001ZHE#04

Quant Time: Feb 24 17:02:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	192922	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	302154	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	290292	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	129384	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	114024	48.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.14%	
35) Dibromofluoromethane	7.57	113	91089	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
50) Toluene-d8	10.08	98	367771	50.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.56%	
62) 4-Bromofluorobenzene	12.40	95	133364	49.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.24%	
Target Compounds						
16) Acetone	3.80	43	4497	4.392	ug/l	98
20) Methylene Chloride	4.53	84	8879	4.346	ug/l	96
43) Isopropyl Acetate	8.14	43	43647	10.095	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022420\
Data File : VN060211.D
Acq On : 24 Feb 2020 15:40
Operator : JC/MD
Sample : PB127001ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127001ZHE#04

Quant Time: Feb 24 17:02:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

