

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022421\
 Data File : VN066001.D
 Acq On : 24 Feb 2021 14:07
 Operator : JC/MD
 Sample : M1473-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GW803-15-20-022321

Quant Time: Feb 25 02:13:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	134244	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	219099	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	202794	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	85554	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.990	65	89239	47.54	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.08%
35) Dibromofluoromethane	7.553	113	67058	49.45	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.90%
50) Toluene-d8	10.058	98	266007	47.79	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.58%
62) 4-Bromofluorobenzene	12.373	95	98868	47.32	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.64%
Target Compounds						
					Qvalue	
16) Acetone	3.791	43	8876	13.53	ug/l	94
30) Chloroform	7.335	83	26203	8.92	ug/l	97
47) Bromodichloromethane	9.370	83	8724	4.14	ug/l	98
60) Dibromochloromethane	10.871	129	1748	1.12	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022421\
Data File : VN066001.D
Acq On : 24 Feb 2021 14:07
Operator : JC/MD
Sample : M1473-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GW803-15-20-022321

Quant Time: Feb 25 02:13:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

