

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022421\
 Data File : VN066005.D
 Acq On : 24 Feb 2021 15:42
 Operator : JC/MD
 Sample : M1473-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GWBR1-119-129-022221

Manual Integrations
 APPROVED

MMDadoda
 2/25/2021 12:15:59 PM

Quant Time: Feb 25 02:14:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.631	168	130912	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	215736	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	193975	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	79449	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	87675	47.89	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.78%		
35) Dibromofluoromethane	7.557	113	64893	48.60	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	97.20%		
50) Toluene-d8	10.058	98	260700	47.57	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	95.14%		
62) 4-Bromofluorobenzene	12.373	95	94362	45.86	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	91.72%		
Target Compounds						Qvalue
12) 1,1-Dichloroethene	3.714	96	1299m	0.88	ug/l	
21) trans-1,2-Dichloroethene	5.007	96	1554m	1.01	ug/l	
27) cis-1,2-Dichloroethene	6.788	96	10169	5.78	ug/l	87
30) Chloroform	7.332	83	8579	2.99	ug/l	92
44) Trichloroethene	8.801	130	32093	19.88	ug/l	99
52) Toluene	10.122	92	62077	15.49	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022421\
Data File : VN066005.D
Acq On : 24 Feb 2021 15:42
Operator : JC/MD
Sample : M1473-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GWBR1-119-129-022221

Manual Integrations
APPROVED

MMDadoda
2/25/2021 12:15:59 PM

Quant Time: Feb 25 02:14:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

