

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022421\
 Data File : VN066006.D
 Acq On : 24 Feb 2021 16:05
 Operator : JC/MD
 Sample : M1473-09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GW803-32-37-022321

Quant Time: Feb 25 02:14:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.631	168	125944	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	204533	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	189892	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	77730	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	83842	47.60	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.20%		
35) Dibromofluoromethane	7.553	113	63226	49.95	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	99.90%		
50) Toluene-d8	10.058	98	250609	48.23	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.46%		
62) 4-Bromofluorobenzene	12.373	95	91891	47.11	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	94.22%		
Target Compounds						Qvalue
12) 1,1-Dichloroethene	3.721	96	4258	3.01	ug/l	96
16) Acetone	3.795	43	5620	8.61	ug/l	96
27) cis-1,2-Dichloroethene	6.795	96	5401	3.19	ug/l	75
30) Chloroform	7.335	83	31751	11.52	ug/l	100
44) Trichloroethene	8.804	130	678	0.44	ug/l	81
47) Bromodichloromethane	9.370	83	10972	5.58	ug/l	94
60) Dibromochloromethane	10.875	129	2183	1.50	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022421\
Data File : VN066006.D
Acq On : 24 Feb 2021 16:05
Operator : JC/MD
Sample : M1473-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GW803-32-37-022321

Quant Time: Feb 25 02:14:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

