

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022421\
 Data File : VN066007.D
 Acq On : 24 Feb 2021 16:29
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Feb 25 02:15:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.631	168	132883	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	218371	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	199252	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	80843	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	88755	47.76	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.52%
35) Dibromofluoromethane	7.553	113	66628	49.30	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.60%
50) Toluene-d8	10.058	98	265044	47.78	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.56%
62) 4-Bromofluorobenzene	12.373	95	96721	46.44	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.88%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022421\
Data File : VN066007.D
Acq On : 24 Feb 2021 16:29
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Feb 25 02:15:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

