

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022421\
 Data File : VN066012.D
 Acq On : 24 Feb 2021 18:27
 Operator : JC/MD
 Sample : M1474-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-TT-VPB-RW6-GW-598-600

Quant Time: Feb 25 02:16:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	119768	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	198841	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	179974	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	73287	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	80206	47.89	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.78%		
35) Dibromofluoromethane	7.553	113	60249	48.96	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	97.92%		
50) Toluene-d8	10.058	98	239525	47.42	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	94.84%		
62) 4-Bromofluorobenzene	12.373	95	87736	46.27	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	92.54%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.853	85	1261	0.80	ug/l	96
16) Acetone	3.792	43	3206	4.52	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022421\
Data File : VN066012.D
Acq On : 24 Feb 2021 18:27
Operator : JC/MD
Sample : M1474-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-TT-VPB-RW6-GW-598-600

Quant Time: Feb 25 02:16:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

