

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022422\
 Data File : VN071059.D
 Acq On : 24 Feb 2022 16:44
 Operator : JC\MD
 Sample : PB142888ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB142888ZHE#05

Quant Time: Feb 25 02:44:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	831014	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1356189	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1226020	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	419279	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	507369	53.448	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.900%
35) Dibromofluoromethane	8.016	113	393252	48.751	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.500%
50) Toluene-d8	10.439	98	1663126	51.339	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.680%
62) 4-Bromofluorobenzene	12.727	95	559502	48.319	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.640%
Target Compounds						
						Qvalue
16) Acetone	4.287	43	41060	7.854	ug/l	99
20) Methylene Chloride	5.093	84	177524	18.684	ug/l	91
43) Isopropyl Acetate	8.557	43	93886	4.202	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022422\
Data File : VN071059.D
Acq On : 24 Feb 2022 16:44
Operator : JC\MD
Sample : PB142888ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB142888ZHE#05

Quant Time: Feb 25 02:44:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

