

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022422\
 Data File : VN071061.D
 Acq On : 24 Feb 2022 17:32
 Operator : JC\MD
 Sample : PB142888ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB142888ZHE#06

Quant Time: Feb 25 02:45:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	757389	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1253215	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1120472	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	383365	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	475993	55.017	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.040%
35) Dibromofluoromethane	8.016	113	361979	48.561	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.120%
50) Toluene-d8	10.439	98	1530218	51.118	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.240%
62) 4-Bromofluorobenzene	12.727	95	519886	48.587	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.180%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	42330	8.884	ug/l	99
18) Methyl Acetate	4.869	43	23256	1.646	ug/l	94
20) Methylene Chloride	5.093	84	177372	20.483	ug/l	94
43) Isopropyl Acetate	8.557	43	146682	7.105	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022422\
Data File : VN071061.D
Acq On : 24 Feb 2022 17:32
Operator : JC\MD
Sample : PB142888ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB142888ZHE#06

Quant Time: Feb 25 02:45:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

