

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022422\
 Data File : VN071048.D
 Acq On : 24 Feb 2022 11:38
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 25 02:41:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	80	0.00
2 T	Dichlorodifluoromethane	0.519	0.509	1.9	81	0.00
3 P	Chloromethane	0.653	0.670	-2.6	89	0.00
4 C	Vinyl Chloride	0.653	0.628	3.8#	81	0.00
5 T	Bromomethane	0.371	0.342	7.8	73	-0.03
6 T	Chloroethane	0.429	0.413	3.7	86	-0.02
7 T	Trichlorofluoromethane	0.819	0.859	-4.9	89	0.00
8 T	Diethyl Ether	0.356	0.362	-1.7	86	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.502	0.501	0.2	84	0.01
10 T	Methyl Iodide	0.696	0.693	0.4	78	0.00
11 T	Tert butyl alcohol	0.104	0.113	-8.7	86	-0.01
12 CM	1,1-Dichloroethene	0.490	0.477	2.7#	83	0.00
13 T	Acrolein	0.093	0.098	-5.4	95	0.00
14 T	Allyl chloride	0.934	0.994	-6.4	90	0.00
15 T	Acrylonitrile	0.307	0.368	-19.9	98	0.00
16 T	Acetone	0.315	0.292	7.3	78	0.00
17 T	Carbon Disulfide	1.404	1.212	13.7	74	0.01
18 T	Methyl Acetate	0.759	0.825	-8.7	99	0.00
19 T	Methyl tert-butyl Ether	1.717	1.867	-8.7	87	0.00
20 T	Methylene Chloride	0.572	0.589	-3.0	87	0.00
21 T	trans-1,2-Dichloroethene	0.525	0.517	1.5	82	0.00
22 T	Diisopropyl ether	1.791	2.114	-18.0	95	0.00
23 T	Vinyl Acetate	1.516	1.779	-17.3	93	0.00
24 P	1,1-Dichloroethane	1.004	1.091	-8.7	90	0.00
25 T	2-Butanone	0.418	0.479	-14.6	91	0.00
26 T	2,2-Dichloropropane	0.834	0.859	-3.0	83	0.00
27 T	cis-1,2-Dichloroethene	0.604	0.628	-4.0	85	0.00
28 T	Bromochloromethane	0.406	0.413	-1.7	81	0.00
29 T	Tetrahydrofuran	0.271	0.335	-23.6	101	0.00
30 C	Chloroform	0.987	1.050	-6.4#	88	0.00
31 T	Cyclohexane	0.988	1.007	-1.9	87	0.00
32 T	1,1,1-Trichloroethane	0.853	0.893	-4.7	84	0.00
33 S	1,2-Dichloroethane-d4	0.571	0.556	2.6	77	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.297	0.280	5.7	73	0.00
36 T	1,1-Dichloropropene	0.487	0.503	-3.3	85	0.00
37 T	Ethyl Acetate	0.523	0.645	-23.3	100	0.00
38 T	Carbon Tetrachloride	0.480	0.496	-3.3	85	0.00
39 T	Methylcyclohexane	0.627	0.620	1.1	80	0.00
40 TM	Benzene	1.441	1.543	-7.1	88	0.00
41 T	Methacrylonitrile	0.275	0.310	-12.7	86	0.00
42 TM	1,2-Dichloroethane	0.491	0.550	-12.0	93	0.00
43 T	Isopropyl Acetate	0.824	0.971	-17.8	94	0.00
44 TM	Trichloroethene	0.380	0.372	2.1	82	0.00
45 C	1,2-Dichloropropane	0.369	0.413	-11.9#	91	0.00
46 T	Dibromomethane	0.243	0.255	-4.9	87	0.00
47 T	Bromodichloromethane	0.491	0.527	-7.3	89	0.00
48 T	Methyl methacrylate	0.390	0.472	-21.0	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022422\
 Data File : VN071048.D
 Acq On : 24 Feb 2022 11:38
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 25 02:41:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.008	-14.3	95	0.00
50 S	Toluene-d8	1.194	1.113	6.8	73	0.00
51 T	4-Methyl-2-Pentanone	0.501	0.626	-25.0	100	0.00
52 CM	Toluene	0.906	0.966	-6.6#	86	0.00
53 T	t-1,3-Dichloropropene	0.534	0.570	-6.7	85	0.00
54 T	cis-1,3-Dichloropropene	0.586	0.634	-8.2	86	0.00
55 T	1,1,2-Trichloroethane	0.354	0.381	-7.6	89	0.00
56 T	Ethyl methacrylate	0.534	0.631	-18.2	91	0.00
57 T	1,3-Dichloropropane	0.613	0.687	-12.1	90	0.00
58 T	2-Chloroethyl Vinyl ether	0.090	0.103	-14.4	88	0.00
59 T	2-Hexanone	0.345	0.431	-24.9	97	0.00
60 T	Dibromochloromethane	0.375	0.400	-6.7	84	0.00
61 T	1,2-Dibromoethane	0.361	0.383	-6.1	86	0.00
62 S	4-Bromofluorobenzene	0.427	0.385	9.8	71	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	81	0.00
64 T	Tetrachloroethene	0.389	0.365	6.2	80	0.00
65 PM	Chlorobenzene	1.077	1.094	-1.6	85	0.00
66 T	1,1,1,2-Tetrachloroethane	0.393	0.406	-3.3	86	0.00
67 C	Ethyl Benzene	1.886	1.996	-5.8#	86	0.00
68 T	m/p-Xylenes	0.728	0.763	-4.8	84	0.00
69 T	o-Xylene	0.712	0.753	-5.8	86	0.00
70 T	Styrene	1.131	1.211	-7.1	84	0.00
71 P	Bromoform	0.313	0.322	-2.9	83	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	73	0.00
73 T	Isopropylbenzene	4.025	4.544	-12.9	85	0.00
74 T	N-amyl acetate	1.286	1.605	-24.8	88	0.00
75 P	1,1,2,2-Tetrachloroethane	1.279	1.476	-15.4	92	0.00
76 T	1,2,3-Trichloropropane	1.177	1.271	-8.0	76	0.00
77 T	Bromobenzene	0.985	1.058	-7.4	82	0.00
78 T	n-propylbenzene	4.522	5.132	-13.5	84	0.00
79 T	2-Chlorotoluene	2.746	3.086	-12.4	85	0.00
80 T	1,3,5-Trimethylbenzene	3.261	3.711	-13.8	84	0.00
81 T	trans-1,4-Dichloro-2-butene	0.363	0.418	-15.2	85	0.00
82 T	4-Chlorotoluene	2.666	2.949	-10.6	82	0.00
83 T	tert-Butylbenzene	3.092	3.336	-7.9	82	0.00
84 T	1,2,4-Trimethylbenzene	3.210	3.567	-11.1	83	0.00
85 T	sec-Butylbenzene	4.074	4.530	-11.2	82	0.00
86 T	p-Isopropyltoluene	3.345	3.679	-10.0	79	0.00
87 T	1,3-Dichlorobenzene	1.744	1.806	-3.6	77	0.00
88 T	1,4-Dichlorobenzene	1.709	1.739	-1.8	77	0.00
89 T	n-Butylbenzene	2.729	2.741	-0.4	73	0.00
90 T	Hexachloroethane	0.597	0.674	-12.9	83	0.00
91 T	1,2-Dichlorobenzene	1.638	1.739	-6.2	80	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.207	0.237	-14.5	83	0.00
93 T	1,2,4-Trichlorobenzene	0.814	0.758	6.9	64	0.00
94 T	Hexachlorobutadiene	0.597	0.519	13.1	67	0.00
95 T	Naphthalene	1.686	1.702	-0.9	64	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022422\
Data File : VN071048.D
Acq On : 24 Feb 2022 11:38
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Feb 25 02:41:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.736	0.715	2.9	66	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6