

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022423\
 Data File : VN076801.D
 Acq On : 24 Feb 2023 11:41
 Operator : JC\MD
 Sample : VN0224WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0224WBL01

Quant Time: Feb 25 01:41:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	268448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.101	114	483739	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	429725	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	168743	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	194495	52.469	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.940%
35) Dibromofluoromethane	8.172	113	150440	47.489	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.980%
50) Toluene-d8	10.572	98	600161	50.729	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.460%
62) 4-Bromofluorobenzene	12.848	95	177075	45.191	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022423\
Data File : VN076801.D
Acq On : 24 Feb 2023 11:41
Operator : JC\MD
Sample : VN0224WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0224WBL01

Quant Time: Feb 25 01:41:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

