

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022423\
 Data File : VN076805.D
 Acq On : 24 Feb 2023 13:26
 Operator : JC\MD
 Sample : 01654-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 UGT

Quant Time: Feb 25 01:44:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	291806	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	520534	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	507292	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	183890	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	210025	52.123	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.240%
35) Dibromofluoromethane	8.172	113	156722	45.975	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.940%
50) Toluene-d8	10.566	98	663927	52.152	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.300%
62) 4-Bromofluorobenzene	12.854	95	205805	48.810	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.620%
Target Compounds						
						Qvalue
16) Acetone	4.437	43	69889	38.830	ug/l #	81
20) Methylene Chloride	5.295	84	17011	4.510	ug/l	88
37) Ethyl Acetate	7.572	43	58663	11.320	ug/l	97
43) Isopropyl Acetate	8.695	43	296674	35.183	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022423\
Data File : VN076805.D
Acq On : 24 Feb 2023 13:26
Operator : JC\MD
Sample : 01654-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
UGT

Quant Time: Feb 25 01:44:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

