

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022423\
 Data File : VN076806.D
 Acq On : 24 Feb 2023 13:50
 Operator : JC\MD
 Sample : 01665-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-231

Quant Time: Feb 25 01:44:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	294585	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	525642	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	501882	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	191188	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	209007	51.381	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.760%		
35) Dibromofluoromethane	8.172	113	159692	46.391	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 92.780%		
50) Toluene-d8	10.572	98	661177	51.431	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.860%		
62) 4-Bromofluorobenzene	12.854	95	205945	48.369	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 96.740%		
Target Compounds						
16) Acetone	4.448	43	101330	55.767	ug/l	Qvalue 93
20) Methylene Chloride	5.290	84	16216	4.258	ug/l #	81
37) Ethyl Acetate	7.560	43	43709	8.353	ug/l #	90
43) Isopropyl Acetate	8.695	43	292051	34.298	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022423\
Data File : VN076806.D
Acq On : 24 Feb 2023 13:50
Operator : JC\MD
Sample : 01665-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-231

Quant Time: Feb 25 01:44:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

