

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022423\
 Data File : VN076808.D
 Acq On : 24 Feb 2023 14:38
 Operator : JC\MD
 Sample : 01639-01
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-05-02212023

Quant Time: Feb 24 14:55:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.225	168	278365	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.101	114	504048	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	408428	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	143694	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.578	65	193219	50.268	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.540%
35) Dibromofluoromethane	8.166	113	148670	45.039	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.080%
50) Toluene-d8	10.566	98	589007	47.780	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.560%
62) 4-Bromofluorobenzene	12.848	95	153166	37.514	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	75.020%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022423\
Data File : VN076808.D
Acq On : 24 Feb 2023 14:38
Operator : JC\MD
Sample : 01639-01
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-05-02212023

Quant Time: Feb 24 14:55:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

