

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022423\
 Data File : VN076799.D
 Acq On : 24 Feb 2023 10:43
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 25 01:40:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	76	0.00
2 T	Dichlorodifluoromethane	0.531	0.484	8.9	67	0.00
3 P	Chloromethane	0.693	0.600	13.4	74	0.00
4 C	Vinyl Chloride	0.682	0.596	12.6#	69	0.00
5 T	Bromomethane	0.413	0.413	0.0	88	0.00
6 T	Chloroethane	0.431	0.443	-2.8	79	0.00
7 T	Trichlorofluoromethane	0.928	0.806	13.1	70	0.00
8 T	Diethyl Ether	0.357	0.389	-9.0	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.545	0.506	7.2	77	0.00
10 T	Methyl Iodide	0.673	0.720	-7.0	73	0.00
11 T	Tert butyl alcohol	0.091	0.105	-15.4	91	-0.01
12 CM	1,1-Dichloroethene	0.516	0.477	7.6#	75	0.00
13 T	Acrolein	0.137	0.156	-13.9	90	0.00
14 T	Allyl chloride	0.894	0.899	-0.6	79	0.00
15 T	Acrylonitrile	0.323	0.369	-14.2	89	0.00
16 T	Acetone	0.308	0.317	-2.9	91	0.00
17 T	Carbon Disulfide	1.363	1.153	15.4	67	0.00
18 T	Methyl Acetate	1.007	1.149	-14.1	90	0.00
19 T	Methyl tert-butyl Ether	1.740	2.056	-18.2	90	0.00
20 T	Methylene Chloride	0.646	0.632	2.2	80	0.00
21 T	trans-1,2-Dichloroethene	0.581	0.540	7.1	77	0.00
22 T	Diisopropyl ether	2.097	2.500	-19.2	90	0.00
23 T	Vinyl Acetate	1.241	1.663	-34.0#	92	0.00
24 P	1,1-Dichloroethane	1.149	1.191	-3.7	84	0.00
25 T	2-Butanone	0.455	0.513	-12.7	89	0.00
26 T	2,2-Dichloropropane	0.829	0.867	-4.6	82	0.00
27 T	cis-1,2-Dichloroethene	0.689	0.721	-4.6	84	0.00
28 T	Bromochloromethane	0.547	0.657	-20.1	94	0.00
29 T	Tetrahydrofuran	0.297	0.342	-15.2	88	0.00
30 C	Chloroform	1.148	1.257	-9.5#	88	0.00
31 T	Cyclohexane	1.083	0.995	8.1	74	0.00
32 T	1,1,1-Trichloroethane	0.966	0.986	-2.1	80	0.00
33 S	1,2-Dichloroethane-d4	0.690	0.780	-13.0	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
35 S	Dibromofluoromethane	0.327	0.324	0.9	87	0.00
36 T	1,1-Dichloropropene	0.473	0.442	6.6	78	0.00
37 T	Ethyl Acetate	0.498	0.565	-13.5	92	0.00
38 T	Carbon Tetrachloride	0.477	0.439	8.0	77	0.00
39 T	Methylcyclohexane	0.538	0.518	3.7	75	0.00
40 TM	Benzene	1.453	1.475	-1.5	84	0.00
41 T	Methacrylonitrile	0.288	0.310	-7.6	86	0.00
42 TM	1,2-Dichloroethane	0.507	0.530	-4.5	88	0.00
43 T	Isopropyl Acetate	0.810	0.945	-16.7	91	0.00
44 TM	Trichloroethene	0.343	0.327	4.7	78	0.00
45 C	1,2-Dichloropropane	0.385	0.403	-4.7#	88	0.00
46 T	Dibromomethane	0.246	0.263	-6.9	89	0.00
47 T	Bromodichloromethane	0.493	0.546	-10.8	92	0.00
48 T	Methyl methacrylate	0.384	0.440	-14.6	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022423\
 Data File : VN076799.D
 Acq On : 24 Feb 2023 10:43
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 25 01:40:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.008	-14.3	97	0.00
50 S	Toluene-d8	1.223	1.273	-4.1	88	0.00
51 T	4-Methyl-2-Pentanone	0.515	0.596	-15.7	93	0.00
52 CM	Toluene	0.877	0.933	-6.4#	86	0.00
53 T	t-1,3-Dichloropropene	0.476	0.576	-21.0	93	0.00
54 T	cis-1,3-Dichloropropene	0.549	0.618	-12.6	89	0.00
55 T	1,1,2-Trichloroethane	0.357	0.389	-9.0	91	0.00
56 T	Ethyl methacrylate	0.502	0.602	-19.9	89	0.00
57 T	1,3-Dichloropropane	0.613	0.685	-11.7	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.230	0.254	-10.4	82	0.00
59 T	2-Hexanone	0.372	0.433	-16.4	91	0.00
60 T	Dibromochloromethane	0.376	0.417	-10.9	89	0.00
61 T	1,2-Dibromoethane	0.357	0.389	-9.0	90	0.00
62 S	4-Bromofluorobenzene	0.405	0.470	-16.0	94	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.315	0.282	10.5	78	0.00
65 PM	Chlorobenzene	1.060	1.049	1.0	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.396	-2.9	88	0.00
67 C	Ethyl Benzene	1.805	1.867	-3.4#	88	0.00
68 T	m/p-Xylenes	0.696	0.736	-5.7	87	0.00
69 T	o-Xylene	0.679	0.732	-7.8	89	0.00
70 T	Styrene	1.097	1.254	-14.3	92	0.00
71 P	Bromoform	0.282	0.318	-12.8	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.665	3.512	4.2	87	0.00
74 T	N-amyl acetate	1.652	1.668	-1.0	90	0.00
75 P	1,1,2,2-Tetrachloroethane	1.327	1.199	9.6	93	0.00
76 T	1,2,3-Trichloropropane	1.011	1.017	-0.6	97	0.00
77 T	Bromobenzene	0.914	0.855	6.5	90	0.00
78 T	n-propylbenzene	4.303	4.219	2.0	88	0.00
79 T	2-Chlorotoluene	2.635	2.500	5.1	90	0.00
80 T	1,3,5-Trimethylbenzene	3.029	3.025	0.1	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.347	0.370	-6.6	94	0.00
82 T	4-Chlorotoluene	2.635	2.500	5.1	90	0.00
83 T	tert-Butylbenzene	2.740	2.666	2.7	87	0.00
84 T	1,2,4-Trimethylbenzene	3.019	3.093	-2.5	91	0.00
85 T	sec-Butylbenzene	3.825	3.794	0.8	86	0.00
86 T	p-Isopropyltoluene	3.055	3.218	-5.3	88	0.00
87 T	1,3-Dichlorobenzene	1.728	1.680	2.8	91	0.00
88 T	1,4-Dichlorobenzene	1.721	1.697	1.4	93	0.00
89 T	n-Butylbenzene	2.657	2.728	-2.7	85	0.00
90 T	Hexachloroethane	0.640	0.595	7.0	89	0.00
91 T	1,2-Dichlorobenzene	1.772	1.719	3.0	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.257	0.227	11.7	91	0.00
93 T	1,2,4-Trichlorobenzene	0.884	0.911	-3.1	88	0.00
94 T	Hexachlorobutadiene	0.433	0.390	9.9	82	0.00
95 T	Naphthalene	2.682	2.926	-9.1	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022423\
Data File : VN076799.D
Acq On : 24 Feb 2023 10:43
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Feb 25 01:40:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.856	0.885	-3.4	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6