

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022423\
 Data File : VN076801.D
 Acq On : 24 Feb 2023 11:41
 Operator : JC\MD
 Sample : VN0224WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0224WBL01

Quant Time: Feb 25 01:41:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	268448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.101	114	483739	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	429725	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	168743	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	194495	52.469	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.940%
35) Dibromofluoromethane	8.172	113	150440	47.489	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.980%
50) Toluene-d8	10.572	98	600161	50.729	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.460%
62) 4-Bromofluorobenzene	12.848	95	177075	45.191	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.380%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022423\
Data File : VN076801.D
Acq On : 24 Feb 2023 11:41
Operator : JC\MD
Sample : VN0224WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0224WBL01

Quant Time: Feb 25 01:41:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.231 min Scan# 1107
Delta R.T. -0.000 min
Lab File: VN076801.D
Acq: 24 Feb 2023 11:41




Instrument :
MSVOA_N
ClientSampleId :
VN0224WBL01


Tgt Ion:168 Resp: 268448
Ion Ratio Lower Upper
168 100
99 57.0 44.8 67.2



#33
1,2-Dichloroethane-d4
Concen: 52.469 ug/l
RT: 8.584 min Scan# 1130
Delta R.T. -0.000 min
Lab File: VN076801.D
Acq: 24 Feb 2023 11:41

Tgt Ion: 65 Resp: 194495
Ion Ratio Lower Upper
65 100
67 54.1 0.0 105.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.101 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN076801.D

Acq: 24 Feb 2023 11:41

Instrument :

MSVOA_N

ClientSampleId :

VN0224WBL01





Tgt Ion:114 Resp: 483739

Ion Ratio Lower Upper

114 100

63 22.0 0.0 40.0

88 13.5 0.0 29.6



#35

Dibromofluoromethane

Concen: 47.489 ug/l

RT: 8.172 min Scan# 1060

Delta R.T. -0.000 min

Lab File: VN076801.D

Acq: 24 Feb 2023 11:41




Tgt Ion:113 Resp: 150440


Ion Ratio Lower Upper

113 100

111 103.9 83.1 124.7

192 17.7 15.6 23.4









#50
Toluene-d8
Concen: 50.729 ug/l
RT: 10.572 min Scan# 1468
Delta R.T. 0.006 min
Lab File: VN076801.D
Acq: 24 Feb 2023 11:41




Instrument :
MSVOA_N
ClientSampleId :
VN0224WBL01


Tgt Ion: 98 Resp: 600161
Ion Ratio Lower Upper
98 100
100 64.0 52.9 79.3



#62
4-Bromofluorobenzene
Concen: 45.191 ug/l
RT: 12.848 min Scan# 1855
Delta R.T. 0.000 min
Lab File: VN076801.D
Acq: 24 Feb 2023 11:41

Tgt Ion: 95 Resp: 177075
Ion Ratio Lower Upper
95 100
174 77.2 0.0 169.2
176 75.5 0.0 163.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.872 min Scan# 1689
Delta R.T. 0.006 min
Lab File: VN076801.D
Acq: 24 Feb 2023 11:41

Instrument :
MSVOA_N
ClientSampleId :
VN0224WBL01

Tgt Ion:117 Resp: 429725
Ion Ratio Lower Upper
117 100
82 55.0 43.4 65.2
119 30.7 26.1 39.1



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.795 min Scan# 2016
Delta R.T. 0.000 min
Lab File: VN076801.D
Acq: 24 Feb 2023 11:41

Tgt Ion:152 Resp: 168743
Ion Ratio Lower Upper
152 100
115 59.4 29.3 87.9
150 157.6 0.0 349.8

