



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022423\
Data File : VN076798.D
Acq On : 24 Feb 2023 09:33
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Title : SW846 8260
Last Update : Tue Feb 14 01:44:47 2023



AutoFind: Scans 1854, 1855, 1856; Background Corrected with Scan 1846

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	22.8	13325	PASS
75	95	30	60	54.9	32141	PASS
95	95	100	100	100.0	58509	PASS
96	95	5	9	6.5	3798	PASS
173	174	0.00	2	1.9	852	PASS
174	95	50	100	78.2	45765	PASS
175	174	5	9	9.0	4107	PASS
176	174	95	101	95.9	43888	PASS
177	176	5	9	6.6	2877	PASS

Average of 12.842 to 12.854 min.: VN076798.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
35.80	133	46.60	110	57.00	2268	66.70	51
36.15	425	47.00	910	57.95	203	67.00	211
36.95	4632	47.85	499	59.90	557	67.95	712
37.95	4259	48.10	63	60.10	224	69.00	746
38.80	412	48.95	3585	60.95	4298	69.85	805
39.00	1412	50.00	13325	62.00	3474	71.70	130
39.90	197	51.05	4742	63.05	3098	72.00	324
44.05	997	52.05	256	63.80	346	72.95	3227
44.90	531	52.80	108	64.10	136	74.00	12578
45.10	351	54.95	400	64.80	199	75.00	32141
46.15	186	55.95	1504	65.00	626	75.95	2586

Average of 12.842 to 12.854 min.: VN076798.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
76.80	85	91.00	269	106.10	83	127.90	144
77.20	378	92.00	2544	106.80	70	129.70	67
77.95	233	92.95	3340	112.80	77	131.00	90
78.85	3628	94.00	7925	115.60	122	134.80	65
79.75	709	95.00	58509	115.90	226	135.90	56
80.90	3598	96.00	3798	116.95	604	139.70	80
81.85	1140	103.10	58	117.80	125	140.70	224
83.05	214	103.70	67	118.00	139	141.00	442
87.05	2756	103.90	361	118.70	371	142.10	64
87.90	3057	104.85	177	119.15	172	142.95	912
90.75	224	105.70	162	127.70	138	145.00	273

Average of 12.842 to 12.854 min.: VN076798.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
145.80	67	173.90	45765				
147.90	115	175.05	4107				
148.70	160	175.90	43888				
152.85	165	176.95	2877				
154.75	172						
161.00	134						
170.60	71						
170.80	202						
171.10	140						
171.85	738						
172.95	852						