

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022425\
 Data File : VN085831.D
 Acq On : 24 Feb 2025 12:39
 Operator : JC\MD
 Sample : VN0224WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0224WBL01

Quant Time: Feb 25 01:51:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	219652	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	415867	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	368217	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	141296	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	171306	60.156	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	120.320%
35) Dibromofluoromethane	8.165	113	149510	54.942	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.880%
50) Toluene-d8	10.565	98	492086	49.702	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.400%
62) 4-Bromofluorobenzene	12.847	95	144556	44.314	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	88.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022425\
Data File : VN085831.D
Acq On : 24 Feb 2025 12:39
Operator : JC\MD
Sample : VN0224WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0224WBL01

Quant Time: Feb 25 01:51:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

