

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022425\
 Data File : VN085839.D
 Acq On : 24 Feb 2025 16:01
 Operator : JC\MD
 Sample : Q1415-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB02222025

Quant Time: Feb 25 01:56:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	226425	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	437095	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	388209	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	145806	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	179257	61.065	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	122.120%	
35) Dibromofluoromethane	8.165	113	152243	53.229	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	106.460%	
50) Toluene-d8	10.565	98	512051	49.207	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	98.420%	
62) 4-Bromofluorobenzene	12.847	95	147244	42.946	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	85.900%	
Target Compounds						
16) Acetone	4.448	43	11136	11.989	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022425\
Data File : VN085839.D
Acq On : 24 Feb 2025 16:01
Operator : JC\MD
Sample : Q1415-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB02222025

Quant Time: Feb 25 01:56:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

