

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022425\
 Data File : VN085840.D
 Acq On : 24 Feb 2025 16:26
 Operator : JC\MD
 Sample : Q1405-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-26-021825

Quant Time: Feb 25 01:56:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	210306	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	401411	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	356777	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	143476	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	169648	62.221	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	124.440%	
35) Dibromofluoromethane	8.165	113	149253	56.823	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	113.640%	
50) Toluene-d8	10.565	98	477627	49.979	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	99.960%	
62) 4-Bromofluorobenzene	12.847	95	139899	44.431	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	88.860%	
Target Compounds						
					Qvalue	
30) Chloroform	7.965	83	5972	1.229	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022425\
Data File : VN085840.D
Acq On : 24 Feb 2025 16:26
Operator : JC\MD
Sample : Q1405-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-26-021825

Quant Time: Feb 25 01:56:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

