

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022425\
 Data File : VN085842.D
 Acq On : 24 Feb 2025 17:14
 Operator : JC\MD
 Sample : Q1405-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-13-021925

Quant Time: Feb 25 01:57:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	206228	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	390930	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	337160	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	124231	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	164335	61.464	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	122.920%
35) Dibromofluoromethane	8.165	113	142571	55.734	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.460%
50) Toluene-d8	10.565	98	464497	49.908	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.820%
62) 4-Bromofluorobenzene	12.847	95	127421	41.553	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	83.100%
Target Compounds						
					Qvalue	
7) Trichlorofluoromethane	3.500	101	12095	2.803	ug/l	98
27) cis-1,2-Dichloroethene	7.488	96	2200	0.779	ug/l	86
44) Trichloroethene	9.347	130	4006	1.406	ug/l	95
64) Tetrachloroethene	11.106	164	6033	2.310	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022425\
Data File : VN085842.D
Acq On : 24 Feb 2025 17:14
Operator : JC\MD
Sample : Q1405-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-13-021925

Quant Time: Feb 25 01:57:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

