

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022519\
 Data File : VN053931.D
 Acq On : 25 Feb 2019 15:56
 Operator : JC/SP
 Sample : VN0225WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0225WBL01

Quant Time: Feb 28 04:28:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 28 04:05:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	629096	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1123528	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1062829	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	395208	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	387488	52.54	ug/l	0.00
Spiked Amount						
			Recovery	=	105.08%	
35) Dibromofluoromethane	7.59	113	360145	49.57	ug/l	0.00
Spiked Amount						
			Recovery	=	99.14%	
50) Toluene-d8	10.09	98	1408729	52.09	ug/l	0.00
Spiked Amount						
			Recovery	=	104.18%	
62) 4-Bromofluorobenzene	12.40	95	461217	51.18	ug/l	0.00
Spiked Amount						
			Recovery	=	102.36%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.81	43	490	Below Cal	#	40

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022519\
Data File : VN053931.D
Acq On : 25 Feb 2019 15:56
Operator : JC/SP
Sample : VN0225WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0225WBL01

Quant Time: Feb 28 04:28:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 28 04:05:45 2019
Response via : Initial Calibration

