

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022519\
 Data File : VN053933.D
 Acq On : 25 Feb 2019 16:56
 Operator : JC/SP
 Sample : PB117322TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB117322TB

Quant Time: Feb 28 04:34:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 28 04:05:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	619935	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1100480	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1071108	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	396493	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	379475	52.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.44%	
35) Dibromofluoromethane	7.59	113	358227	50.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.68%	
50) Toluene-d8	10.09	98	1401721	52.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.82%	
62) 4-Bromofluorobenzene	12.40	95	460264	52.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.30%	
Target Compounds						
20) Methylene Chloride	4.55	84	15640	2.175	ug/l	94
95) Naphthalene	15.13	128	7443	3.655	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022519\
Data File : VN053933.D
Acq On : 25 Feb 2019 16:56
Operator : JC/SP
Sample : PB117322TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB117322TB

Quant Time: Feb 28 04:34:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 28 04:05:45 2019
Response via : Initial Calibration

