

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022519\
 Data File : VN053935.D
 Acq On : 25 Feb 2019 17:49
 Operator : JC/SP
 Sample : K1348-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampled :
 OK-03-022219-B

Quant Time: Feb 28 04:36:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 28 04:05:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	673014	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1165353	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1132646	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	420151	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	407348	51.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.26%	
35) Dibromofluoromethane	7.59	113	379385	50.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.68%	
50) Toluene-d8	10.09	98	1480667	52.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.56%	
62) 4-Bromofluorobenzene	12.40	95	486699	52.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.14%	
Target Compounds						
16) Acetone	3.82	43	26010	0.311	ug/l	99
20) Methylene Chloride	4.55	84	113918	14.590	ug/l	91
95) Naphthalene	15.13	128	8512	3.691	ug/l #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022519\
 Data File : VN053935.D
 Acq On : 25 Feb 2019 17:49
 Operator : JC/SP
 Sample : K1348-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 Client Sampled :
 OK-03-022219-B

Quant Time: Feb 28 04:36:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 28 04:05:45 2019
 Response via : Initial Calibration

