

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022519\
 Data File : VN053939.D
 Acq On : 25 Feb 2019 19:36
 Operator : JC/SP
 Sample : K1592-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-4

Quant Time: Feb 28 04:59:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 28 04:05:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	619149	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1143202	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1080631	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	382321	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	400917	55.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.48%	
35) Dibromofluoromethane	7.59	113	372963	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
50) Toluene-d8	10.09	98	1439742	52.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.64%	
62) 4-Bromofluorobenzene	12.40	95	453247	49.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.86%	
Target Compounds						
20) Methylene Chloride	4.55	84	18612	2.591	ug/l	98
43) Isopropyl Acetate	8.17	43	38014	2.966	ug/l #	89
95) Naphthalene	15.14	128	6795	3.632	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022519\
Data File : VN053939.D
Acq On : 25 Feb 2019 19:36
Operator : JC/SP
Sample : K1592-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-4

Quant Time: Feb 28 04:59:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 28 04:05:45 2019
Response via : Initial Calibration

