

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022520\
 Data File : VN060221.D
 Acq On : 25 Feb 2020 9:56
 Operator : JC/MD
 Sample : VN0225WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0225WBL01

Quant Time: Feb 25 16:00:09 2020
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	171785	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	268652	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	246322	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	101271	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	101168	48.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.80%	
35) Dibromofluoromethane	7.57	113	81508	49.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.72%	
50) Toluene-d8	10.08	98	322295	49.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.12%	
62) 4-Bromofluorobenzene	12.40	95	110499	46.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.48%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022520\
Data File : VN060221.D
Acq On : 25 Feb 2020 9:56
Operator : JC/MD
Sample : VN0225WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0225WBL01

Quant Time: Feb 25 16:00:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

