

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022520\
 Data File : VN060224.D
 Acq On : 25 Feb 2020 11:26
 Operator : JC/MD
 Sample : PB127055TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127055TB

Quant Time: Feb 25 16:06:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	166884	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	259454	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	242288	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	106378	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	97371	47.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.90%	
35) Dibromofluoromethane	7.57	113	78808	49.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.82%	
50) Toluene-d8	10.08	98	314277	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
62) 4-Bromofluorobenzene	12.40	95	111159	48.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.34%	
Target Compounds						
16) Acetone	3.80	43	4562	5.150	ug/l	100
20) Methylene Chloride	4.53	84	7899	4.470	ug/l	96
43) Isopropyl Acetate	8.14	43	43754	11.785	ug/l #	89
95) Naphthalene	15.13	128	7665	1.414	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022520\
Data File : VN060224.D
Acq On : 25 Feb 2020 11:26
Operator : JC/MD
Sample : PB127055TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127055TB

Quant Time: Feb 25 16:06:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

