

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022520\
 Data File : VN060225.D
 Acq On : 25 Feb 2020 11:49
 Operator : JC/MD
 Sample : PB127069TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127069TB

Quant Time: Feb 25 16:16:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	175761	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	277029	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	257761	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	113136	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	103279	48.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.58%	
35) Dibromofluoromethane	7.57	113	82865	48.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.32%	
50) Toluene-d8	10.08	98	331718	49.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.94%	
62) 4-Bromofluorobenzene	12.40	95	119329	48.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.86%	
Target Compounds						
16) Acetone	3.79	43	4215	4.518	ug/l	91
20) Methylene Chloride	4.53	84	7982	4.289	ug/l	97
43) Isopropyl Acetate	8.14	43	43970	11.092	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022520\
Data File : VN060225.D
Acq On : 25 Feb 2020 11:49
Operator : JC/MD
Sample : PB127069TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127069TB

Quant Time: Feb 25 16:16:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

