

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022520\
 Data File : VN060228.D
 Acq On : 25 Feb 2020 12:57
 Operator : JC/MD
 Sample : L1668-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30830

Quant Time: Feb 25 16:42:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	177320	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	276765	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	267053	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	127825	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	104195	48.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.58%	
35) Dibromofluoromethane	7.57	113	84900	49.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.80%	
50) Toluene-d8	10.08	98	339111	50.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.24%	
62) 4-Bromofluorobenzene	12.40	95	125090	50.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.62%	
Target Compounds						
					Qvalue	
16) Acetone	3.79	43	40105	42.612	ug/l	100
20) Methylene Chloride	4.53	84	7429	3.956	ug/l	96
25) 2-Butanone	6.81	43	17541	14.819	ug/l	98
43) Isopropyl Acetate	8.14	43	40089	10.122	ug/l #	90
51) 4-Methyl-2-Pentanone	9.97	43	32233	14.061	ug/l	100
52) Toluene	10.14	92	20441	4.192	ug/l	96
95) Naphthalene	15.13	128	30356	4.661	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022520\
Data File : VN060228.D
Acq On : 25 Feb 2020 12:57
Operator : JC/MD
Sample : L1668-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30830

Quant Time: Feb 25 16:42:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

