

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022520\
 Data File : VN060234.D
 Acq On : 25 Feb 2020 15:13
 Operator : JC/MD
 Sample : PB127055ZHE#15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127055ZHE#15

Quant Time: Feb 25 22:49:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	181734	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	285940	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	274841	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	125346	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	107423	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
35) Dibromofluoromethane	7.57	113	86793	49.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.76%	
50) Toluene-d8	10.08	98	352706	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
62) 4-Bromofluorobenzene	12.40	95	127925	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
Target Compounds						
16) Acetone	3.80	43	6069	6.292	ug/l	99
20) Methylene Chloride	4.52	84	8199	4.260	ug/l #	94
43) Isopropyl Acetate	8.14	43	43885	10.725	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022520\
Data File : VN060234.D
Acq On : 25 Feb 2020 15:13
Operator : JC/MD
Sample : PB127055ZHE#15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127055ZHE#15

Quant Time: Feb 25 22:49:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

