

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022520\
 Data File : VN060242.D
 Acq On : 25 Feb 2020 18:14
 Operator : JC/MD
 Sample : L1666-02DL 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NB-30871DL

Quant Time: Feb 25 23:13:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	229146	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	366347	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	344809	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	177746	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	139545	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
35) Dibromofluoromethane	7.57	113	107389	47.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.38%	
50) Toluene-d8	10.08	98	443134	49.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.94%	
62) 4-Bromofluorobenzene	12.40	95	176613	54.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.40%	
Target Compounds						
					Qvalue	
16) Acetone	3.78	43	81322	66.863	ug/l	100
40) Benzene	8.02	78	19348	1.922	ug/l	99
52) Toluene	10.14	92	14836	2.298	ug/l	100
68) m/p-Xylenes	11.61	106	7243	1.523	ug/l	96
95) Naphthalene	15.13	128	19329	2.134	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022520\
Data File : VN060242.D
Acq On : 25 Feb 2020 18:14
Operator : JC/MD
Sample : L1666-02DL 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NB-30871DL

Quant Time: Feb 25 23:13:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

