

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022521\
 Data File : VN066024.D
 Acq On : 25 Feb 2021 13:46
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Feb 26 01:19:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	197916	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	331444	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	296924	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	120880	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	117843	42.58	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	85.16%
35) Dibromofluoromethane	7.550	113	95655	46.63	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.26%
50) Toluene-d8	10.055	98	404506	48.04	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.08%
62) 4-Bromofluorobenzene	12.373	95	140880	44.57	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.14%
Target Compounds						
95) Naphthalene	15.100	128	11165	4.55	ug/l	Qvalue # 70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022521\
Data File : VN066024.D
Acq On : 25 Feb 2021 13:46
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Feb 26 01:19:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

