

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022521\
 Data File : VN066027.D
 Acq On : 25 Feb 2021 14:57
 Operator : JC/MD
 Sample : M1484-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GW804-10-15-022421

Quant Time: Feb 26 01:20:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	170995	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	285821	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	254482	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	101926	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	103449	43.26	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	86.52%		
35) Dibromofluoromethane	7.553	113	84642	47.85	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	95.70%		
50) Toluene-d8	10.058	98	347446	47.85	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	95.70%		
62) 4-Bromofluorobenzene	12.373	95	120922	44.36	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	88.72%		
Target Compounds					Qvalue	
16) Acetone	3.798	43	3352	2.88	ug/l #	88
30) Chloroform	7.328	83	1250	0.33	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022521\
Data File : VN066027.D
Acq On : 25 Feb 2021 14:57
Operator : JC/MD
Sample : M1484-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GW804-10-15-022421

Quant Time: Feb 26 01:20:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

