

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022521\
 Data File : VN066029.D
 Acq On : 25 Feb 2021 15:44
 Operator : JC/MD
 Sample : M1484-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GW804-34-39-022421

Quant Time: Feb 26 01:21:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	153978	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	251962	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	228036	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	91652	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	96747	44.93	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	89.86%		
35) Dibromofluoromethane	7.553	113	74409	47.72	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	95.44%		
50) Toluene-d8	10.058	98	308017	48.12	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.24%		
62) 4-Bromofluorobenzene	12.373	95	107696	44.82	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	89.64%		
Target Compounds					Qvalue	
16) Acetone	3.795	43	3710	3.91	ug/l	97
30) Chloroform	7.332	83	44642	13.24	ug/l	100
47) Bromodichloromethane	9.367	83	16780	6.93	ug/l #	96
60) Dibromochloromethane	10.875	129	3274	1.83	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022521\
Data File : VN066029.D
Acq On : 25 Feb 2021 15:44
Operator : JC/MD
Sample : M1484-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GW804-34-39-022421

Quant Time: Feb 26 01:21:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

