

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022521\
 Data File : VN066032.D
 Acq On : 25 Feb 2021 16:55
 Operator : JC/MD
 Sample : PB134803TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134803TB

Quant Time: Feb 26 01:22:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	147123	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	240084	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	217420	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	89939	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	94282	45.83	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	91.66%		
35) Dibromofluoromethane	7.553	113	72193	48.59	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	97.18%		
50) Toluene-d8	10.058	98	293130	48.06	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.12%		
62) 4-Bromofluorobenzene	12.373	95	106882	46.68	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	93.36%		
Target Compounds					Qvalue	
16) Acetone	3.795	43	6068	7.84	ug/l	97
20) Methylene Chloride	4.521	84	5735	3.01	ug/l	83
43) Isopropyl Acetate	8.126	43	10623	3.12	ug/l #	89
95) Naphthalene	15.103	128	5235	4.00	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022521\
Data File : VN066032.D
Acq On : 25 Feb 2021 16:55
Operator : JC/MD
Sample : PB134803TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134803TB

Quant Time: Feb 26 01:22:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

