

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022521\
 Data File : VN066033.D
 Acq On : 25 Feb 2021 17:18
 Operator : JC/MD
 Sample : PB134803ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134803ZHE#01

Quant Time: Feb 26 01:23:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	134492	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	218675	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	199399	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	82428	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	86014	45.73	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	91.46%		
35) Dibromofluoromethane	7.553	113	65528	48.42	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	96.84%		
50) Toluene-d8	10.058	98	266939	48.05	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.10%		
62) 4-Bromofluorobenzene	12.373	95	97287	46.65	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	93.30%		
Target Compounds					Qvalue	
16) Acetone	3.791	43	6063	8.71	ug/l	91
20) Methylene Chloride	4.515	84	5324	3.05	ug/l #	82
43) Isopropyl Acetate	8.129	43	10341	3.33	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022521\
Data File : VN066033.D
Acq On : 25 Feb 2021 17:18
Operator : JC/MD
Sample : PB134803ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134803ZHE#01

Quant Time: Feb 26 01:23:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

