

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022521\
 Data File : VN066035.D
 Acq On : 25 Feb 2021 18:06
 Operator : JC/MD
 Sample : PB134803ZHE#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134803ZHE#03

Quant Time: Feb 26 01:24:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.631	168	126338	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	206088	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	188504	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	79319	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	83946	47.51	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.02%
35) Dibromofluoromethane	7.553	113	62782	49.22	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.44%
50) Toluene-d8	10.058	98	250376	47.82	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.64%
62) 4-Bromofluorobenzene	12.373	95	92161	46.89	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.78%
Target Compounds						
					Qvalue	
16) Acetone	3.792	43	5461	8.29	ug/l	96
20) Methylene Chloride	4.525	84	4980	3.04	ug/l #	84
43) Isopropyl Acetate	8.126	43	9735	3.33	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022521\
Data File : VN066035.D
Acq On : 25 Feb 2021 18:06
Operator : JC/MD
Sample : PB134803ZHE#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134803ZHE#03

Quant Time: Feb 26 01:24:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

