

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022521\
 Data File : VN066038.D
 Acq On : 25 Feb 2021 19:40
 Operator : JC/MD
 Sample : M1476-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ML-CON

Quant Time: Feb 26 01:25:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	116703	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	192596	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	176459	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	71936	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	78187	47.91	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.82%
35) Dibromofluoromethane	7.553	113	58180	48.81	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.62%
50) Toluene-d8	10.058	98	232109	47.44	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.88%
62) 4-Bromofluorobenzene	12.373	95	84821	46.18	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.36%
Target Compounds						
					Qvalue	
16) Acetone	3.791	43	9289	16.62	ug/l	96
20) Methylene Chloride	4.531	84	4971	3.28	ug/l	92
43) Isopropyl Acetate	8.126	43	6884	2.52	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022521\
Data File : VN066038.D
Acq On : 25 Feb 2021 19:40
Operator : JC/MD
Sample : M1476-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ML-CON

Quant Time: Feb 26 01:25:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

