

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022522\
 Data File : VN071074.D
 Acq On : 25 Feb 2022 16:58
 Operator : JC\MD
 Sample : N1664-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Quant Time: Feb 26 03:43:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	856912	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1389661	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1254266	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	443785	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	532055	54.354	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.700%	
35) Dibromofluoromethane	8.016	113	402632	48.711	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 97.420%	
50) Toluene-d8	10.439	98	1702326	51.284	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.560%	
62) 4-Bromofluorobenzene	12.727	95	595497	50.189	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 100.380%	
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	58495	10.851	ug/l	95
20) Methylene Chloride	5.098	84	113680	11.603	ug/l	95
43) Isopropyl Acetate	8.557	43	106864	4.668	ug/l #	86
93) 1,2,4-Trichlorobenzene	15.263	180	3514	2.412	ug/l	98
95) Naphthalene	15.504	128	12350	5.550	ug/l	96
96) 1,2,3-Trichlorobenzene	15.698	180	3951	3.014	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022522\
Data File : VN071074.D
Acq On : 25 Feb 2022 16:58
Operator : JC\MD
Sample : N1664-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Quant Time: Feb 26 03:43:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

